

Nolato Treff Quality

CleanRoom Pure[®]

Basic Quality and Purity Standard of all TreffLab Products

**“Absolutely PURE
Right From The Start”**



Statement

Nolato Treff AG confirms that TreffLab consumables and products marked with the CleanRoom[®] Pure quality symbol have been manufactured in microbiologically controlled clean rooms class 8, ISO 14644-1. To avoid direct human contact, employees are fully protective dressed and masked.

Quality Parameters

Quality standard for original sealed packages

- Free of RNase (< 5x10E-9 Kunitz units)
- Free of DNase (< 5x10E-6 Kunitz units)
- Free of DNA (< 2 pg DNA)
- Free of PCR inhibitors (< 10 copies)
- Free of endotoxins (< 0.001 EU/ml)
- Free of R and ATP (< 30 RLU)

Analyses are performed by an independent Swiss laboratory for molecular biological & microbiological analytics, accredited ISO/IEC17025 for real-time PCR.

06. March 2024



Richard Wiesli
Sales & Marketing Manager,
TreffLab



Peter Klose
Head Quality Management